

Instructions For Use, IFU Prosthetists

Xtend® Connect A100



Lindhe Xtend AB
Olofsdalsvägen 40 A, 302 41 Halmstad, Sweden
info@lindhe.se
Phone: +46 (0)35 262 92 00
www.lindhextend.com

SIZE AND WEIGHT GUIDE

The maximum weight for Xtend Connect is 150 kg.
Always check that the patient's weight is correct. The maximum weight may never be exceeded.
In case components with different weight limits are used in the prosthetic leg, it is the component with the lowest weight limit that never shall be exceeded.
The max weight is limited to 100 kg for patients who do high-intensive sports activities.

INTENDED USE:

XTEND CONNECT is a prosthetic component intended for use by above-knee or lower leg amputees who want to remove or change prosthetic parts.
XTEND CONNECT is designed to be fitted between two prosthetic parts in a prosthetic leg and designed to fit standard connections according to common industry practice.

PATIENT INDICATIONS

The PATIENT is physically disabled and the disability can vary from amputation of a lower limb to hip amputation. Double amputees can also use the product.
More specifically, the product is suitable for the following types of applications:

- Patients who want to switch between different prosthetic feet depending on the type of activity. For example, a sports foot and everyday foot.
- Above-knee amputees who want to switch between different prosthetic knees depending on the activity. For example, a waterproof prosthetic and everyday knee.
- Patients who want to switch between indoor and outdoor prosthetics.
- Patients who find changing or taking off their shoes difficult and time-consuming.
- Patients who wear tight trousers and find it difficult to put them on over rigid prosthetic feet.
- Patients who find it difficult to get in and out of cramped vehicles and need to remove parts of their prosthesis to sit comfortably during the journey.


CONTRAINDICATIONS

The product is unsuitable for the following types of applications:

- Patients who weigh over 150 kg cannot use Xtend Connect. The permitted maximum weight is 150 kg.
- Patients with a diagnosis of dementia, e.g. Alzheimer's, should not use Xtend Connect.
- Patients who lift heavy weights and weigh around 150 kg should not use Xtend Connect. The permitted max. weight incl. external load is 150 kg.
- The product is not suitable for patients who do extreme sports.
- The product is not suitable for patients exceeding 100 kg who do high-intensive sports activities.

XTEND® CONNECT AND ITS DIFFERENT COMPONENTS

Xtend Connect consists of two components:
Quick fit unit – can be opened and closed with the aid of a locking wheel and a safety push-button
Base unit – fits into the quick fit unit.



Item number	Description
The parts are sold separately.	
Quick fit units	
A100-Q30-150	Xtend Connect quick fit unit with tube clamp for 30 mm tube
A100-Q4-150	Xtend Connect quick fit unit with 4-hole adapter
A100-QF-150	Xtend Connect quick fit unit with female pyramid adapter
A100-QP-150	Xtend Connect Quick fit unit with M36 Prong adapter
Base units	
A100-BF-150	Xtend Connect base unit with female pyramid adapter
A100-BM-150	Xtend Connect base unit with male pyramid adapter
A100-B30-150	Xtend Connect base unit with 30 mm tube

GENERAL INSTRUCTIONS:

- Review the care, safety and patient instructions with the patient at the time of fitting the product. The patient must be given the special folder for patient guidance.
- Get the patient to practise connecting and removing Xtend Connect several times until he/she has mastered the process. This is extremely important to ensure that product satisfaction is high.
- ⚠ It is VERY IMPORTANT that the patient is seated while removing and connecting Xtend Connect. The patient may NEVER stand while connecting and removing Xtend Connect.
- Under no circumstances may the patient walk on Xtend Connect if it is not correctly locked. This can cause the patient to fall and get injured.
- Only authorised prosthetist may test the product on patients.
- Xtend Connect can be fitted on a number of different connecting parts and in a number of different configurations. Consider carefully how it will be used before ordering.
- Before ordering components, it is important to have analysed whether the patient is suited to using this type of quick fit connector and if he/she is able to follow the instructions.
- Before mounting the Xtend Connect on the patient, assure that the joining components fits well to the interfaces of the Xtend Connect. Test-lock the complete unit mounted on the prosthetic leg before fitting it to the patient. Only external components using industry standard may be used.
- All adverse events that occurs and are related to this product should be reported to the manufacturer and the local authority in the country of occurred event.
- After finished life cycle, the product shall be disposed in metal recycling.
- ⚠ During installation of Xtend Connect tube coupling, do not cut the tube shorter than the length needed, being certain to factor in the portion of the tube that must be fully inserted into the receiving tube clamp.

INSTALLATION

Only trained personnel such as an prosthetist may install, carry out maintenance on and fit the product on patients/users.
Xtend Connect can be fitted with the quick fit unit facing upwards and the base unit facing downwards, as well as vice versa. It is the connecting components in the prosthetic leg that determine which is most suitable. Remember that right-handed people will want the locking wheel on their right and left-handers vice versa.

WHAT YOU NEED FOR INSTALLATION

- Torque wrench 4 mm Allen key
- Thread locker e.g. Loctite 243
- Thread locker e.g. Loctite 275

INSTALLATION OF TUBE CLAMP COUPLING

	Tightening torque	Thread locker
Clamp screw	7 Nm	E.g. Loctite 243

Make sure the tube goes all the way into Xtend Connect. Xtend Connect is designed to fit flush with the tube. For this reason, it may sometimes be necessary to grind the tube so the parts fit. When the tube is all the way in and everything is stable, tighten the clamp screw with the torque wrench.
When everything has been installed, use Loctite to lock the screw. This must be done before the patient leaves the clinic.



INSTALLATION OF 4 HOLE ADAPTER

	Tightening torque	Thread locker
Adjusting screws	–	E.g. Loctite 243

Tighten all the four screws. When the installation is ready, use Loctite to secure all four screws. This must be done before the patient leaves the workshop.
It is possible to order different length of the screws from Lindhe Xtend if this should be required.



INSTALLATION OF FEMALE PYRAMID

	Tightening torque	Thread locker
Adjusting screws	15 Nm	E.g. Loctite 243

Tighten the adjustment screws using the torque wrench once you have tested the product on the patient.
When everything has been installed, use Loctite to lock each screw one at a time. This must be done before the patient leaves the clinic.




INSTALLATION OF M36 PRONG ADAPTER

	Tightening torque	Thread locker
Locking thread on M36	–	E.g. Loctite 275

Screw the Quick Fit unit into the female prong adapter all the way in. If needed, you can unscrew a slight bit in order to get a correct alignment of the locking wheel but this may never exceed 360°.

When everything is installed, use Loctite to lock the prong adapter. This must be done before the patient walks on the prosthesis.



INSTALLATION OF TUBE

	Tightening torque	Thread locker
Tube	–	–

Cut the tube to the appropriate length. Be cautious of sharp edges after cutting. Do not cut the tube shorter than the length needed, being certain to factor in the portion of the tube that must be fully inserted into the receiving tube clamp.

It may be helpful to roughen up the surface portion of the tube that fits into the clamp to increase grip at components interface.

Sufficient protection equipment shall be used when cutting and grinding the tube.

Insert the tube into the receiving component. Ensure that the tube is completely inserted and makes contact with bottom of tube clamp.

SETTINGS AND FINAL FITTING

When you have fitted the different parts, try opening and closing Xtend Connect several times before you test it on the patient. Listen for strange noises or rotation in connections and Xtend Connect itself.

When you are sure that everything is working properly, you can try it out on the patient and make the final adjustments to heights and settings.

Get the patient to walk forwards and backwards and listen carefully to make sure there is no noise.

Also ask the patient to take a few big steps to the side and then turn around. There should not be any visible rotation in Xtend Connect.

REGULAR MAINTENANCE

We recommend having Xtend Connect inspected by a prosthetist every six months.

With very active use, shorter inspection interval may be necessary.

CLEANING INSTRUCTIONS

You should clean all Xtend Connect’s parts at least once a week and after use in salt water or sand.

We recommend using warm water and a gentle cleaning agent. Do as follows:

1. Remove your prosthetic leg. *Tip: Sit on a chair to clean it.*
2. Separate the Xtend Connect parts.
3. Clean all Xtend Connect parts. They should remain attached to their respective prosthetic leg parts. You can either lower them into a warm bath or hold them under warm running water. Scrub slightly with a mild cleaning agent.
Tip: It may be helpful to use cotton buds.
4. Open and close the locking wheel a couple of times until the wedge runs smoothly and the locking wheel can be turned easily.
5. Rinse carefully.
6. Wipe with a lint-free dry cloth or paper so that the surfaces are clean and dry.

TIP

If the purpose is to be able to switch prosthetic components, install these components AFTER you have adjusted your first complete prosthetic leg. This makes it easier to make minor adjustments to the respective prosthetic component and saves you time.

CHANGING FEET:

Because different feet have different heights, remember to fit Xtend Connect as high as possible up the tube that connects to the prosthetic foot. By doing this, you can adjust for different prosthetic foot heights by truncating the connecting tube to the correct height. It is also easier for the prosthetic user to change the foot if there is a tube to hold onto.



CHANGING THE KNEE JOINT:

Bear in mind that changing knee joints can be heavy work for the prosthesis user, and that both hands are often needed to perform the change.

Adjustments are very important, and easiest to achieve if you use the female pyramid as a connecting component to the knee joint.

It is helpful to rotate Xtend Connect so that the quick fit unit faces downwards. This does not affect the functioning of the component, and can sometimes be easier for the patient to manage.



CHANGING TROUSERS:

Here you can put Xtend Connect wherever you like, but the lower down the prosthetic foot the connection is, the easier it is to pull on tight trousers.



TROUBLESHOOTING

The locking wheel seizes.


Cause: The locking wheel has dirt in the turning wheel.

Solution: Clean Xtend Connect. Blow it clean with compressed air or use warm water and a gentle cleaning agent. Then rotate the locking wheel 10 turns in an unobstructed position.

The locking wheel seizes despite cleaning according to instructions.

Cause: A coating of salt water or chloride water might have remained in the locking wheel mechanism.

Solution: Add one drop of lubrication inside the groove of Locking wheel. We recommend that you use silicon lubrication that not attract dirt or dust.



It is not possible to separate the parts

Cause: Dirt between the connecting surfaces.

Solution: First clean the area according to the instructions in the patient manual. Use water or compressed air. Carefully try to separate the parts. If this does not work, use more force and try to separate the parts. If necessary, use some lubricant, e.g. 5-56 to separate the parts more easily. Once the parts have been separated, clean them carefully.

If this does not work, contact Lindhe Xtend.

The locking wheel does not pop out properly and the parts therefore cannot be separated

Cause: Dirt in the mechanism of the locking function

Solution: Clean Xtend Connect, preferably using compressed air. In stages, you may need to clean, try to screw the locking wheel and pull a little, clean it again and pull some more. If the locking wheel is springy and pops straight back, it is in its outer position.

The safety push-button does not pop out to the locking position.

Cause: Dirt in the mechanism

Solution: Take Xtend Connect apart and clean it according to the cleaning instructions in the patient manual.

If you cannot get the safety push-button to function, the patient must not use the product. Contact Lindhe Xtend to discuss a solution.

⚠ WARNINGS

⚠ WARNINGS TO THE PATIENT

- Xtend Connect may not be connected or disconnected while the patient is standing. This can cause the patient to fall.
- Sand and dirt can make it difficult to fit and remove Xtend Connect, e.g. for use at the beach. It might also cause abnormal wear on the device.
- Be careful not to get your fingers caught when fitting Xtend Connect. Do not hold onto the connection surfaces between the two parts.
- Do not walk on Xtend Connect if the parts are not correctly fitted and correctly locked. This can cause you to fall and injure yourself.
- If you are in a warm environment, watch out for warm parts on Xtend Connect.

⚠ OTHER WARNINGS

- Xtend Connect is intended for single use only. Do not re-use an Xtend Connect that has been used on another patient since its use and history are unknown.
- Xtend Connect is manufactured with several standardised interfaces to fit all common connection components. It is the prosthetist's responsibility to correctly adapt this coupling to other prosthesis components.
- The prosthetic component is NOT to be altered or modified outside the specifications of the product by the user or prosthetist.
- If, on inspection, the parts look worn, they must be replaced as soon as possible by a qualified prosthetist.
- Inspection must be performed if the device is subjected to external violence.
- During installation of Xtend Connect tube coupling, be cautious of sharp edges after cutting the tube.

GUARANTEE

12-month guarantee from the delivery date from Lindhe Xtend.

Failure to follow the user instructions will invalidate the guarantee. Failure to follow the weight recommendations or exposing the component to unreasonable loads such as heavy lifting or similar activities which may otherwise have injured a human body part, will invalidate the terms of the guarantee.

Xtend Connect A100 is covered by the following patents:

- Germany DE 20 2018 000 520
- Patent pending in EU, USA and Japan